

Radiofrequency wire for the recanalization of central vein occlusions that have failed conventional endovascular techniques

Marcelo Guimaraes, Claudio Schonholz, Christopher Hannegan, Michael Bret Anderson, June Shi, Bayne Selby Jr.

Radiofrequency Wire for the Recanalization of Central Vein Occlusions that Have Failed Conventional Endovascular Techniques - Journal of Vascular and Interventional Radiology.

INTRODUCTION

This single center retrospective review evaluated the feasibility and safety of using the PowerWire® Radiofrequency (RF) guidewire to recanalize central venous occlusions (CVOs) in cases where conventional endovascular recanalization had failed.

METHODS

Forty-three benign, chronic, but symptomatic CVOs were treated with RF wire puncture and subsequent stent implantation in 42 patients.

- CVO recanalization was first attempted with a combination of hydrophilic guide wires and semi-curved catheters. Only after an inability to cross the lesion was PowerWire RF guidewire used.
- Six subclavian, 29 brachiocephalic, and 8 superior vena cava CVOs were recanalized with the PowerWire RF guidewire.*

Technical success: Successful crossing of the obstruction using PowerWire RF guidewire with subsequent balloon angioplasty and stent placement.

Clinical success: The absence of symptoms and presence of functional arteriovenous shunts in patients who had surgical dialysis access at the time of follow-up after RF wire recanalization.

RESULTS

- All 42 patients (100%) had successful recanalization facilitated by the PowerWire RF guidewire, with all 43 lesions (100%) achieving technical success.
- Follow-up showed that 40 of 42 patients (95.2%) had patent stents at 6 months, and clinical success at 9 months.
- Unrelated to the use of RF wire puncture, there was one adverse event where a patient experienced cardiac tamponade during balloon angioplasty requiring a percutaneous drain catheter.

DISCUSSION

- The PowerWire RF guidewire can be used to recanalize symptomatic and chronic CVOs when traditional endovascular techniques have failed.
- The RF wire should only be advanced a few millimeters during each RF application and targeted towards a simple one loop snare at the opposite side of the venous occlusion. Alignment of the RF wire tip and snare should be verified in anteroposterior, RAO, and LAO views.

**Disclaimer: this study includes both on-label and off-label use of the PowerWire® RF Guidewire. Before use, consult product labels and Instructions for Use for Indications for Use, Contraindications, Warnings, Precautions, Adverse Events and Directions for Use.*

PRM-00877 EN J-1,2,3. V-1 © Copyright Baylis Medical Technologies Inc., 2023. PowerWire and the Baylis Medical Technologies logo are trademarks and/or registered trademarks of Baylis Medical Technologies Inc. in the USA and/or other countries. Other trademarks are property of their respective owners. Baylis Medical Technologies reserves the right to change specifications or to incorporate design changes without notice and without incurring any obligation relating to equipment previously manufactured or delivered. CAUTION: Federal Law (USA) restricts the sale of these devices to or by the order of a physician. Before use, consult product labels and Instructions for Use for Indications for Use, Contraindications, Warnings, Precautions, Adverse Events and Directions for Use. Patents pending and/or issued.

Products shown may not be approved or available for sale in all jurisdictions.

Baylis
MED TECH

Baylis Medical Technologies Inc
2645 Matheson Blvd East
Mississauga, ON Canada L4W 5S4
Tel: 1 (888)-505-4885
www.baylismedtech.com
info@baylismedtech.com